

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024743.D
 Acq On : 12 Oct 2021 04:57
 Operator : JC/MD
 Sample : M4158-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-3-BOTTOM-GRAB

Quant Time: Oct 12 08:10:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236653	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102415	54.204	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.400%
35) Dibromofluoromethane	5.397	113	80357	51.113	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.220%
50) Toluene-d8	8.653	98	303179	52.396	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.800%
62) 4-Bromofluorobenzene	11.085	95	110288	48.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.960%
Target Compounds						Qvalue
16) Acetone	2.386	43	17175	11.885	ug/l	91
20) Methylene Chloride	2.794	84	7025	3.915	ug/l	92
43) Isopropyl Acetate	6.348	43	15874	3.686	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024743.D
Acq On : 12 Oct 2021 04:57
Operator : JC/MD
Sample : M4158-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-3-BOTTOM-GRAB

Quant Time: Oct 12 08:10:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

