

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024744.D
 Acq On : 12 Oct 2021 05:20
 Operator : JC/MD
 Sample : M4158-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-4-BOTTOM-GRAB

Quant Time: Oct 12 08:11:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228468	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99500	54.854	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.700%	
35) Dibromofluoromethane	5.397	113	77478	51.047	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.100%	
50) Toluene-d8	8.653	98	288001	51.556	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.120%	
62) 4-Bromofluorobenzene	11.085	95	110991	50.533	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.060%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	17257	12.932	ug/l	97
20) Methylene Chloride	2.794	84	7214	4.269	ug/l	96
43) Isopropyl Acetate	6.348	43	15539	3.738	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024744.D
Acq On : 12 Oct 2021 05:20
Operator : JC/MD
Sample : M4158-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-4-BOTTOM-GRAB

Quant Time: Oct 12 08:11:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

