

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024745.D
 Acq On : 12 Oct 2021 05:44
 Operator : JC/MD
 Sample : M4158-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T8-1-TOP-GRAB

Quant Time: Oct 12 08:11:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229902	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98665	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100980	55.090	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.180%
35) Dibromofluoromethane	5.397	113	78456	51.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.740%
50) Toluene-d8	8.653	98	293495	52.212	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.420%
62) 4-Bromofluorobenzene	11.085	95	108328	49.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.020%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	24595	22.576	ug/l	93
20) Methylene Chloride	2.794	84	6853	3.943	ug/l	96
43) Isopropyl Acetate	6.355	43	16263	3.888	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024745.D
Acq On : 12 Oct 2021 05:44
Operator : JC/MD
Sample : M4158-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T8-1-TOP-GRAB

Quant Time: Oct 12 08:11:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

