

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024746.D
 Acq On : 12 Oct 2021 06:07
 Operator : JC/MD
 Sample : M4158-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T8-2-TOP-GRAB

Quant Time: Oct 12 08:11:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	223901	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99043	54.408	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.820%	
35) Dibromofluoromethane	5.391	113	77254	51.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.880%	
50) Toluene-d8	8.653	98	285775	52.201	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.400%	
62) 4-Bromofluorobenzene	11.085	95	105822	49.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.320%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	28320	27.861	ug/l #	86
20) Methylene Chloride	2.794	84	6557	3.756	ug/l	94
43) Isopropyl Acetate	6.354	43	15392	3.778	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024746.D
Acq On : 12 Oct 2021 06:07
Operator : JC/MD
Sample : M4158-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T8-2-TOP-GRAB

Quant Time: Oct 12 08:11:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

