

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024750.D
 Acq On : 12 Oct 2021 07:41
 Operator : JC/MD
 Sample : M4158-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T8-2-BOTTOM-GRAB

Quant Time: Oct 12 08:12:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227708	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102165	56.376	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.760%
35) Dibromofluoromethane	5.391	113	79679	52.673	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.340%
50) Toluene-d8	8.653	98	294326	52.864	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.720%
62) 4-Bromofluorobenzene	11.085	95	109796	50.156	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.320%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	26540	25.608	ug/l	# 86
20) Methylene Chloride	2.794	84	6757	3.929	ug/l	87
43) Isopropyl Acetate	6.348	43	16033	3.870	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024750.D
Acq On : 12 Oct 2021 07:41
Operator : JC/MD
Sample : M4158-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T8-2-BOTTOM-GRAB

Quant Time: Oct 12 08:12:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

