

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101121\
 Data File : VX024729.D
 Acq On : 11 Oct 2021 22:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 12 06:55:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	50.000	61.896	-23.8	101	0.00
3 P	Chloromethane	50.000	53.323	-6.6	90	0.00
4 C	Vinyl Chloride	50.000	54.595	-9.2#	91	0.00
5 T	Bromomethane	50.000	67.373	-34.7#	112	0.00
6 T	Chloroethane	50.000	50.784	-1.6	89	0.00
7 T	Trichlorofluoromethane	50.000	52.552	-5.1	89	0.00
8 T	Diethyl Ether	50.000	51.993	-4.0	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.067	-4.1	89	0.00
10 T	Methyl Iodide	50.000	48.988	2.0	82	0.00
11 T	Tert butyl alcohol	250.000	211.660	15.3	71	0.00
12 CM	1,1-Dichloroethene	50.000	53.267	-6.5#	92	0.00
13 T	Acrolein	250.000	282.425	-13.0	96	0.00
14 T	Allyl chloride	50.000	51.168	-2.3	86	0.00
15 T	Acrylonitrile	250.000	245.753	1.7	84	0.00
16 T	Acetone	250.000	284.269	-13.7	92	0.00
17 T	Carbon Disulfide	50.000	50.574	-1.1	89	0.00
18 T	Methyl Acetate	50.000	51.748	-3.5	91	0.00
19 T	Methyl tert-butyl Ether	50.000	52.393	-4.8	90	0.00
20 T	Methylene Chloride	50.000	55.554	-11.1	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.373	-4.7	91	0.00
22 T	Diisopropyl ether	50.000	52.665	-5.3	90	0.00
23 T	Vinyl Acetate	250.000	269.381	-7.8	89	0.00
24 P	1,1-Dichloroethane	50.000	52.702	-5.4	90	0.00
25 T	2-Butanone	250.000	276.579	-10.6	93	0.00
26 T	2,2-Dichloropropane	50.000	45.288	9.4	77	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.310	-4.6	91	0.00
28 T	Bromochloromethane	50.000	50.707	-1.4	95	0.00
29 T	Tetrahydrofuran	250.000	247.014	1.2	82	0.00
30 C	Chloroform	50.000	53.232	-6.5#	92	0.00
31 T	Cyclohexane	50.000	51.020	-2.0	86	0.00
32 T	1,1,1-Trichloroethane	50.000	54.463	-8.9	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.321	1.4	86	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	86	0.00
35 S	Dibromofluoromethane	50.000	50.943	-1.9	87	0.00
36 T	1,1-Dichloropropene	50.000	53.656	-7.3	92	0.00
37 T	Ethyl Acetate	50.000	49.177	1.6	85	0.00
38 T	Carbon Tetrachloride	50.000	54.582	-9.2	89	0.00
39 T	Methylcyclohexane	50.000	53.124	-6.2	87	0.00
40 TM	Benzene	50.000	53.906	-7.8	93	0.00
41 T	Methacrylonitrile	50.000	51.260	-2.5	86	0.00
42 TM	1,2-Dichloroethane	50.000	54.274	-8.5	91	0.00
43 T	Isopropyl Acetate	50.000	52.674	-5.3	90	0.00
44 TM	Trichloroethene	50.000	54.127	-8.3	94	0.00
45 C	1,2-Dichloropropane	50.000	53.600	-7.2#	92	0.00
46 T	Dibromomethane	50.000	52.767	-5.5	91	0.00
47 T	Bromodichloromethane	50.000	55.732	-11.5	91	0.00
48 T	Methyl methacrylate	50.000	58.339	-16.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024729.D
 Acq On : 11 Oct 2021 22:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 12 06:55:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	943.008	5.7	76	0.00
50 S	Toluene-d8	50.000	50.014	-0.0	85	0.00
51 T	4-Methyl-2-Pentanone	250.000	293.520	-17.4	97	0.00
52 CM	Toluene	50.000	54.843	-9.7#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	53.034	-6.1	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.855	-9.7	90	0.00
55 T	1,1,2-Trichloroethane	50.000	56.118	-12.2	94	0.00
56 T	Ethyl methacrylate	50.000	48.666	2.7	80	0.00
57 T	1,3-Dichloropropane	50.000	53.961	-7.9	92	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	263.457	-5.4	85	0.00
59 T	2-Hexanone	250.000	290.204	-16.1	94	0.00
60 T	Dibromochloromethane	50.000	51.461	-2.9	91	0.00
61 T	1,2-Dibromoethane	50.000	55.628	-11.3	93	0.00
62 S	4-Bromofluorobenzene	50.000	51.254	-2.5	88	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	87	0.00
64 T	Tetrachloroethene	50.000	59.661	-19.3	101	0.00
65 PM	Chlorobenzene	50.000	54.292	-8.6	92	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	56.118	-12.2	94	0.00
67 C	Ethyl Benzene	50.000	55.472	-10.9#	91	0.00
68 T	m/p-Xylenes	100.000	111.318	-11.3	90	0.00
69 T	o-Xylene	50.000	53.681	-7.4	89	0.00
70 T	Styrene	50.000	52.514	-5.0	86	0.00
71 P	Bromoform	50.000	48.474	3.1	90	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	86	0.00
73 T	Isopropylbenzene	50.000	54.493	-9.0	91	0.00
74 T	N-ethyl acetate	50.000	52.501	-5.0	88	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.443	-2.9	89	0.00
76 T	1,2,3-Trichloropropane	50.000	52.410	-4.8	90	0.00
77 T	Bromobenzene	50.000	53.001	-6.0	92	0.00
78 T	n-propylbenzene	50.000	52.747	-5.5	88	0.00
79 T	2-Chlorotoluene	50.000	52.441	-4.9	89	0.00
80 T	1,3,5-Trimethylbenzene	50.000	55.576	-11.2	90	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.472	5.1	83	0.00
82 T	4-Chlorotoluene	50.000	52.969	-5.9	89	0.00
83 T	tert-Butylbenzene	50.000	53.221	-6.4	88	0.00
84 T	1,2,4-Trimethylbenzene	50.000	57.141	-14.3	93	0.00
85 T	sec-Butylbenzene	50.000	51.301	-2.6	84	0.00
86 T	p-Isopropyltoluene	50.000	55.570	-11.1	90	0.00
87 T	1,3-Dichlorobenzene	50.000	50.355	-0.7	87	0.00
88 T	1,4-Dichlorobenzene	50.000	49.735	0.5	88	0.00
89 T	n-Butylbenzene	50.000	54.650	-9.3	89	0.00
90 T	Hexachloroethane	50.000	53.565	-7.1	90	0.00
91 T	1,2-Dichlorobenzene	50.000	51.126	-2.3	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.172	-0.3	82	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.472	-2.9	88	0.00
94 T	Hexachlorobutadiene	50.000	49.926	0.1	85	0.00
95 T	Naphthalene	50.000	56.371	-12.7	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024729.D
Acq On : 11 Oct 2021 22:19
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 12 06:55:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.977	-6.0	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6