



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024709.D
 Acq On : 11 Oct 2021 10:42
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Title : SW846 8260
 Last Update : Thu Sep 23 18:24:34 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	7683	PASS
75	95	30	60	52.4	20801	PASS
95	95	100	100	100.0	39669	PASS
96	95	5	9	6.7	2676	PASS
173	174	0.00	2	1.2	411	PASS
174	95	50	100	83.6	33144	PASS
175	174	5	9	7.8	2599	PASS
176	174	95	101	98.7	32717	PASS
177	176	5	9	6.3	2054	PASS