

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032029.D
 Acq On : 11 Oct 2022 12:09
 Operator : JC/MD
 Sample : VX1011WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1011WBL01

Quant Time: Oct 11 16:22:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	69807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	103340	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69668	48.325	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.640%
35) Dibromofluoromethane	5.391	113	56013	47.885	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.760%
50) Toluene-d8	8.647	98	206947	49.166	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.340%
62) 4-Bromofluorobenzene	11.079	95	68118	45.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032029.D
Acq On : 11 Oct 2022 12:09
Operator : JC/MD
Sample : VX1011WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1011WBL01

Quant Time: Oct 11 16:22:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

