

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032033.D
 Acq On : 11 Oct 2022 13:49
 Operator : JC/MD
 Sample : N5046-10
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R-10

Quant Time: Oct 11 16:24:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	79350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	115419	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60784	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79273	48.374	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.740%
35) Dibromofluoromethane	5.391	113	58180	44.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.060%
50) Toluene-d8	8.647	98	231831	49.314	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.620%
62) 4-Bromofluorobenzene	11.085	95	67293	40.489	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.980%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032033.D
Acq On : 11 Oct 2022 13:49
Operator : JC/MD
Sample : N5046-10
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R-10

Quant Time: Oct 11 16:24:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

