

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032035.D
 Acq On : 11 Oct 2022 14:36
 Operator : JC/MD
 Sample : N5091-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-101022

Quant Time: Oct 11 16:25:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	112057	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76054	49.550	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.100%
35) Dibromofluoromethane	5.385	113	59640	47.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.040%
50) Toluene-d8	8.653	98	225929	49.500	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.000%
62) 4-Bromofluorobenzene	11.079	95	73754	45.708	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032035.D
Acq On : 11 Oct 2022 14:36
Operator : JC/MD
Sample : N5091-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-101022

Quant Time: Oct 11 16:25:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

