

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032036.D
 Acq On : 11 Oct 2022 14:59
 Operator : JC/MD
 Sample : N4937-04DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-109DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/12/2022
 Supervised By : Mahesh Dadoda 10/14/2022

Quant Time: Oct 12 05:12:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	107465	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64471	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73572	49.026	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.060%		
35) Dibromofluoromethane	5.385	113	57662	47.402	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.800%		
50) Toluene-d8	8.647	98	218027	49.810	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.620%		
62) 4-Bromofluorobenzene	11.079	95	71694	46.330	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.660%		
Target Compounds						Qvalue
11) Tert butyl alcohol	3.056	59	10754m	57.597	ug/l	
19) Methyl tert-butyl Ether	3.117	73	153058	41.209	ug/l	99
40) Benzene	6.037	78	4370	0.925	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032036.D
Acq On : 11 Oct 2022 14:59
Operator : JC/MD
Sample : N4937-04DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-109DL

Quant Time: Oct 12 05:12:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/12/2022
Supervised By : Mahesh Dadoda 10/14/2022

