

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032040.D
 Acq On : 11 Oct 2022 16:32
 Operator : JC/MD
 Sample : N4937-12 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-118D

Quant Time: Oct 12 05:14:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	69652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	104216	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70404	48.944	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.880%
35) Dibromofluoromethane	5.385	113	57330	48.599	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.200%
50) Toluene-d8	8.653	98	210315	49.546	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	68000	45.313	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.620%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	5076	28.362	ug/l #	73
19) Methyl tert-butyl Ether	3.111	73	201131	56.494	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032040.D
Acq On : 11 Oct 2022 16:32
Operator : JC/MD
Sample : N4937-12 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-118D

Quant Time: Oct 12 05:14:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

