

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032045.D
 Acq On : 11 Oct 2022 18:28
 Operator : JC/MD
 Sample : N5100-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-TB-20221007

Quant Time: Oct 12 05:17:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72400	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	107158	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74216	49.636	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.280%
35) Dibromofluoromethane	5.385	113	58816	48.489	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.980%
50) Toluene-d8	8.653	98	217936	49.932	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.860%
62) 4-Bromofluorobenzene	11.079	95	71925	46.613	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%
Target Compounds						
30) Chloroform	5.093	83	3937	1.672	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032045.D
Acq On : 11 Oct 2022 18:28
Operator : JC/MD
Sample : N5100-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-TB-20221007

Quant Time: Oct 12 05:17:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

