

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032046.D
 Acq On : 11 Oct 2022 18:51
 Operator : JC/MD
 Sample : N5100-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-EB-20221010

Quant Time: Oct 12 05:17:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72324	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	109194	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75611	50.622	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.240%
35) Dibromofluoromethane	5.385	113	59384	48.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	8.647	98	218771	49.188	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.380%
62) 4-Bromofluorobenzene	11.079	95	73367	46.661	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.320%
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	3730	1.585	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032046.D
Acq On : 11 Oct 2022 18:51
Operator : JC/MD
Sample : N5100-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-EB-20221010

Quant Time: Oct 12 05:17:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

