

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032058.D
 Acq On : 12 Oct 2022 03:55
 Operator : JC/MD
 Sample : N5066-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-1(10-11)

Quant Time: Oct 12 06:22:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75962	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	116364	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81997	52.268	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.540%
35) Dibromofluoromethane	5.385	113	63788	48.428	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.860%
50) Toluene-d8	8.647	98	242035	51.066	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.140%
62) 4-Bromofluorobenzene	11.079	95	82248	49.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.180%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	22675	42.529	ug/l	100
20) Methylene Chloride	2.788	84	7376	5.265	ug/l	88
37) Ethyl Acetate	4.715	43	13991m	8.508	ug/l	
43) Isopropyl Acetate	6.342	43	58830	23.580	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032058.D
Acq On : 12 Oct 2022 03:55
Operator : JC/MD
Sample : N5066-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-1(10-11)

Quant Time: Oct 12 06:22:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

