

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032060.D
 Acq On : 12 Oct 2022 04:42
 Operator : JC/MD
 Sample : N5067-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B6(5-6)

Quant Time: Oct 12 06:23:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79970	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121190	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84153	50.954	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.900%
35) Dibromofluoromethane	5.385	113	65420	47.689	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.380%
50) Toluene-d8	8.653	98	251619	50.974	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.079	95	85382	48.927	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.860%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	23282	41.479	ug/l	96
20) Methylene Chloride	2.794	84	7854	5.325	ug/l	89
25) 2-Butanone	4.587	43	3227	3.678	ug/l	91
37) Ethyl Acetate	4.721	43	14178	8.279	ug/l	98
43) Isopropyl Acetate	6.348	43	61449	23.649	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032060.D
Acq On : 12 Oct 2022 04:42
Operator : JC/MD
Sample : N5067-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B6(5-6)

Quant Time: Oct 12 06:23:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

