

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032063.D
 Acq On : 12 Oct 2022 05:51
 Operator : JC/MD
 Sample : N5067-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B6(22-23)

Quant Time: Oct 12 06:24:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118330	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83661	51.810	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.620%
35) Dibromofluoromethane	5.385	113	63302	47.260	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.520%
50) Toluene-d8	8.647	98	246348	51.112	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.220%
62) 4-Bromofluorobenzene	11.079	95	83514	49.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.020%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	22672	41.312	ug/l	99
37) Ethyl Acetate	4.721	43	13233	7.914	ug/l	98
43) Isopropyl Acetate	6.342	43	60530	23.858	ug/l	98
95) Naphthalene	13.774	128	132653	24.127	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032063.D
Acq On : 12 Oct 2022 05:51
Operator : JC/MD
Sample : N5067-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B6(22-23)

Quant Time: Oct 12 06:24:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

