

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032065.D
 Acq On : 12 Oct 2022 06:38
 Operator : JC/MD
 Sample : N5067-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B3(23-24)

Quant Time: Oct 12 13:25:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77790	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82680	51.465	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.940%
35) Dibromofluoromethane	5.385	113	64758	48.424	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.647	98	247561	51.445	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.880%
62) 4-Bromofluorobenzene	11.079	95	84118	49.445	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.900%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	25306	46.348	ug/l	93
20) Methylene Chloride	2.794	84	8688	6.055	ug/l	89
25) 2-Butanone	4.574	43	4304	5.043	ug/l #	88
37) Ethyl Acetate	4.721	43	13122	7.860	ug/l	98
43) Isopropyl Acetate	6.342	43	57631	22.751	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032065.D
Acq On : 12 Oct 2022 06:38
Operator : JC/MD
Sample : N5067-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B3(23-24)

Quant Time: Oct 12 13:25:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

