

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
 Data File : VX032070.D
 Acq On : 12 Oct 2022 08:34
 Operator : JC/MD
 Sample : N5067-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B4(13-14)

Quant Time: Oct 12 13:27:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	117203	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83129	52.625	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.260%
35) Dibromofluoromethane	5.385	113	63734	48.041	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	8.647	98	244341	51.184	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.360%
62) 4-Bromofluorobenzene	11.079	95	80496	47.696	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.400%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	23590	43.941	ug/l	98
20) Methylene Chloride	2.788	84	7757	5.498	ug/l #	93
37) Ethyl Acetate	4.714	43	13058	7.884	ug/l	98
43) Isopropyl Acetate	6.336	43	58447	23.259	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101122\
Data File : VX032070.D
Acq On : 12 Oct 2022 08:34
Operator : JC/MD
Sample : N5067-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B4(13-14)

Quant Time: Oct 12 13:27:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

