

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
 Data File : VX038224.D
 Acq On : 11 Oct 2023 13:13
 Operator : JC/MD
 Sample : 04788-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PMW-1

Quant Time: Oct 12 02:43:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197502	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81868	45.571	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.140%
35) Dibromofluoromethane	5.385	113	65452	45.992	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.980%
50) Toluene-d8	8.647	98	245210	45.091	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.180%#
62) 4-Bromofluorobenzene	11.079	95	92526	44.880	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
Data File : VX038224.D
Acq On : 11 Oct 2023 13:13
Operator : JC/MD
Sample : 04788-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PMW-1

Quant Time: Oct 12 02:43:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

