

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
 Data File : VX038230.D
 Acq On : 11 Oct 2023 15:32
 Operator : JC/MD
 Sample : 04803-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-7-10-13

Quant Time: Oct 12 02:46:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108026	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101561	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78799	45.351	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.700%		
35) Dibromofluoromethane	5.385	113	63592	46.075	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.160%		
50) Toluene-d8	8.653	98	243961	46.257	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.520%		
62) 4-Bromofluorobenzene	11.079	95	92705	46.366	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.740%		
Target Compounds						
30) Chloroform	5.092	83	4898	1.992	ug/l	93
52) Toluene	8.720	92	47827	13.793	ug/l	97
64) Tetrachloroethene	9.275	164	861	0.677	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
Data File : VX038230.D
Acq On : 11 Oct 2023 15:32
Operator : JC/MD
Sample : 04803-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-7-10-13

Quant Time: Oct 12 02:46:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

