

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
 Data File : VX038234.D
 Acq On : 11 Oct 2023 17:05
 Operator : JC/MD
 Sample : 04803-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-8-23-26

Quant Time: Oct 12 02:47:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189641	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98691	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77058	45.550	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.100%
35) Dibromofluoromethane	5.385	113	61961	45.343	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.680%
50) Toluene-d8	8.647	98	239651	45.895	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.800%#
62) 4-Bromofluorobenzene	11.079	95	88906	44.911	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.820%
Target Compounds						
16) Acetone	2.380	43	1633	2.313	ug/l	98
30) Chloroform	5.105	83	1953	0.816	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
Data File : VX038234.D
Acq On : 11 Oct 2023 17:05
Operator : JC/MD
Sample : 04803-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-8-23-26

Quant Time: Oct 12 02:47:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

