

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
 Data File : VX038237.D
 Acq On : 11 Oct 2023 18:15
 Operator : JC/MD
 Sample : 04803-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-9-12-15

Quant Time: Oct 12 02:48:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	183169	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77181	47.029	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.060%
35) Dibromofluoromethane	5.385	113	61040	46.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.500%
50) Toluene-d8	8.647	98	235741	46.742	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.480%
62) 4-Bromofluorobenzene	11.079	95	89902	47.019	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.040%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2022	2.952	ug/l	90
44) Trichloroethene	7.129	130	1876	1.409	ug/l	68
64) Tetrachloroethene	9.275	164	1940	1.556	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
Data File : VX038237.D
Acq On : 11 Oct 2023 18:15
Operator : JC/MD
Sample : 04803-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-9-12-15

Quant Time: Oct 12 02:48:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

