

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
 Data File : VX038238.D
 Acq On : 11 Oct 2023 18:38
 Operator : JC/MD
 Sample : 04803-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-9-23-26

Quant Time: Oct 12 02:49:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105457	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189772	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80123	47.237	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.480%
35) Dibromofluoromethane	5.391	113	64332	47.046	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.100%
50) Toluene-d8	8.647	98	241882	46.291	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.580%
62) 4-Bromofluorobenzene	11.079	95	90641	45.756	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.520%
Target Compounds						Qvalue
64) Tetrachloroethene	9.275	164	1286	1.030	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101123\
Data File : VX038238.D
Acq On : 11 Oct 2023 18:38
Operator : JC/MD
Sample : 04803-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-9-23-26

Quant Time: Oct 12 02:49:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

