

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
 Data File : VX043357.D
 Acq On : 11 Oct 2024 10:55
 Operator : JC/MD
 Sample : VX1011MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1011MBL01

Quant Time: Oct 14 02:42:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	81837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147555	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53943	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	67373	50.409	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.820%
35) Dibromofluoromethane	5.385	113	49246	48.394	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	8.647	98	177332	50.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.079	95	62689	48.932	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
Data File : VX043357.D
Acq On : 11 Oct 2024 10:55
Operator : JC/MD
Sample : VX1011MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1011MBL01

Quant Time: Oct 14 02:42:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

