

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
 Data File : VX043365.D
 Acq On : 11 Oct 2024 14:04
 Operator : JC/MD
 Sample : P4388-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Oct 14 02:47:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67560	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	126634	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48828	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60369	54.714	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.420%
35) Dibromofluoromethane	5.391	113	41779	47.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.680%
50) Toluene-d8	8.647	98	147896	48.868	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.079	95	54953	49.980	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
Data File : VX043365.D
Acq On : 11 Oct 2024 14:04
Operator : JC/MD
Sample : P4388-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Oct 14 02:47:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

