

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
 Data File : VX043367.D
 Acq On : 11 Oct 2024 16:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 14 02:48:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	50.000	52.959	-5.9	91	0.00
3 P	Chloromethane	50.000	50.346	-0.7	89	0.00
4 C	Vinyl Chloride	50.000	52.785	-5.6#	93	0.00
5 T	Bromomethane	50.000	52.478	-5.0	95	0.00
6 T	Chloroethane	50.000	57.611	-15.2	101	0.00
7 T	Trichlorofluoromethane	50.000	56.664	-13.3	103	0.00
8 T	Diethyl Ether	50.000	54.469	-8.9	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	54.161	-8.3	96	0.00
10 T	Methyl Iodide	50.000	53.274	-6.5	92	0.00
11 T	Tert butyl alcohol	250.000	279.724	-11.9	96	0.00
12 CM	1,1-Dichloroethene	50.000	51.002	-2.0#	91	0.00
13 T	Acrolein	250.000	217.523	13.0	77	0.00
14 T	Allyl chloride	50.000	52.935	-5.9	93	0.00
15 T	Acrylonitrile	250.000	264.006	-5.6	93	0.00
16 T	Acetone	250.000	262.480	-5.0	100	0.00
17 T	Carbon Disulfide	50.000	46.296	7.4	84	0.00
18 T	Methyl Acetate	50.000	48.889	2.2	83	0.00
19 T	Methyl tert-butyl Ether	50.000	54.694	-9.4	96	0.00
20 T	Methylene Chloride	50.000	50.343	-0.7	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	53.435	-6.9	95	0.00
22 T	Diisopropyl ether	50.000	54.355	-8.7	96	0.00
23 T	Vinyl Acetate	250.000	219.603	12.2	72	0.00
24 P	1,1-Dichloroethane	50.000	54.513	-9.0	96	0.00
25 T	2-Butanone	250.000	268.982	-7.6	94	0.00
26 T	2,2-Dichloropropane	50.000	53.279	-6.6	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.474	-6.9	95	0.00
28 T	Bromochloromethane	50.000	45.091	9.8	83	0.00
29 T	Tetrahydrofuran	250.000	261.298	-4.5	94	0.00
30 C	Chloroform	50.000	53.283	-6.6#	96	-0.01
31 T	Cyclohexane	50.000	52.893	-5.8	96	0.00
32 T	1,1,1-Trichloroethane	50.000	51.999	-4.0	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.556	-7.1	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	87	0.00
35 S	Dibromofluoromethane	50.000	53.488	-7.0	94	0.00
36 T	1,1-Dichloropropene	50.000	53.562	-7.1	94	0.00
37 T	Ethyl Acetate	50.000	53.561	-7.1	94	0.00
38 T	Carbon Tetrachloride	50.000	53.283	-6.6	91	0.00
39 T	Methylcyclohexane	50.000	54.720	-9.4	95	0.00
40 TM	Benzene	50.000	53.542	-7.1	94	0.00
41 T	Methacrylonitrile	50.000	55.281	-10.6	91	0.00
42 TM	1,2-Dichloroethane	50.000	55.987	-12.0	96	0.00
43 T	Isopropyl Acetate	50.000	53.783	-7.6	91	0.00
44 TM	Trichloroethene	50.000	51.759	-3.5	92	0.00
45 C	1,2-Dichloropropane	50.000	53.437	-6.9#	92	0.00
46 T	Dibromomethane	50.000	54.544	-9.1	94	0.00
47 T	Bromodichloromethane	50.000	53.530	-7.1	90	0.00
48 T	Methyl methacrylate	50.000	52.862	-5.7	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101124\
 Data File : VX043367.D
 Acq On : 11 Oct 2024 16:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 14 02:48:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	1,4-Dioxane	1000.000	1024.144	-2.4	86	0.00
50 S	Toluene-d8	50.000	50.335	-0.7	86	0.00
51 T	4-Methyl-2-Pentanone	250.000	265.987	-6.4	90	0.00
52 CM	Toluene	50.000	52.343	-4.7#	89	0.00
53 T	t-1,3-Dichloropropene	50.000	54.905	-9.8	90	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.895	-7.8	91	0.00
55 T	1,1,2-Trichloroethane	50.000	53.258	-6.5	90	0.00
56 T	Ethyl methacrylate	50.000	54.601	-9.2	88	0.00
57 T	1,3-Dichloropropane	50.000	53.168	-6.3	90	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	263.242	-5.3	84	0.00
59 T	2-Hexanone	250.000	266.330	-6.5	89	0.00
60 T	Dibromochloromethane	50.000	51.774	-3.5	85	0.00
61 T	1,2-Dibromoethane	50.000	53.431	-6.9	90	0.00
62 S	4-Bromofluorobenzene	50.000	52.581	-5.2	88	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	84	0.00
64 T	Tetrachloroethene	50.000	53.028	-6.1	90	0.00
65 PM	Chlorobenzene	50.000	51.731	-3.5	89	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.791	-5.6	87	0.00
67 C	Ethyl Benzene	50.000	52.568	-5.1#	87	0.00
68 T	m/p-Xylenes	100.000	103.690	-3.7	87	0.00
69 T	o-Xylene	50.000	51.395	-2.8	87	0.00
70 T	Styrene	50.000	53.465	-6.9	87	0.00
71 P	Bromoform	50.000	51.969	-3.9	84	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	85	0.00
73 T	Isopropylbenzene	50.000	50.854	-1.7	86	0.00
74 T	N-ethyl acetate	50.000	51.987	-4.0	88	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.144	-2.3	88	0.00
76 T	1,2,3-Trichloropropane	50.000	52.088	-4.2	91	0.00
77 T	Bromobenzene	50.000	50.361	-0.7	87	0.00
78 T	n-propylbenzene	50.000	53.073	-6.1	87	0.00
79 T	2-Chlorotoluene	50.000	50.869	-1.7	89	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.612	-3.2	88	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.614	-1.2	88	0.00
82 T	4-Chlorotoluene	50.000	51.654	-3.3	89	0.00
83 T	tert-Butylbenzene	50.000	50.164	-0.3	86	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.757	-3.5	88	0.00
85 T	sec-Butylbenzene	50.000	51.391	-2.8	86	0.00
86 T	p-Isopropyltoluene	50.000	52.216	-4.4	88	0.00
87 T	1,3-Dichlorobenzene	50.000	51.357	-2.7	89	0.00
88 T	1,4-Dichlorobenzene	50.000	50.197	-0.4	89	0.00
89 T	n-Butylbenzene	50.000	54.483	-9.0	89	0.00
90 T	Hexachloroethane	50.000	50.140	-0.3	83	0.00
91 T	1,2-Dichlorobenzene	50.000	51.160	-2.3	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.637	0.7	83	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.147	-6.3	88	0.00
94 T	Hexachlorobutadiene	50.000	54.085	-8.2	89	0.00
95 T	Naphthalene	50.000	50.680	-1.4	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
Data File : VX043367.D
Acq On : 11 Oct 2024 16:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 14 02:48:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.634	-5.3	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6