

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
 Data File : VX043357.D
 Acq On : 11 Oct 2024 10:55
 Operator : JC/MD
 Sample : VX1011MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1011MBL01

Quant Time: Oct 14 02:42:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	81837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147555	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53943	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	67373	50.409	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.820%
35) Dibromofluoromethane	5.385	113	49246	48.394	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	8.647	98	177332	50.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.079	95	62689	48.932	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.860%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101124\
Data File : VX043357.D
Acq On : 11 Oct 2024 10:55
Operator : JC/MD
Sample : VX1011MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1011MBL01

Quant Time: Oct 14 02:42:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX043357.D
Acq: 11 Oct 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :
VX1011MBL01


Tgt Ion:168 Resp: 81837
Ion Ratio Lower Upper
168 100
99 60.1 49.7 74.5



#33
1,2-Dichloroethane-d4
Concen: 50.409 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.006 min
Lab File: VX043357.D
Acq: 11 Oct 2024 10:55

Tgt Ion: 65 Resp: 67373
Ion Ratio Lower Upper
65 100
67 48.8 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX043357.D
Acq: 11 Oct 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :
VX1011MBL01


Tgt Ion:114 Resp: 147555
Ion Ratio Lower Upper
114 100
63 21.7 0.0 43.0
88 16.3 0.0 33.0



#35
Dibromofluoromethane
Concen: 48.394 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX043357.D
Acq: 11 Oct 2024 10:55

Tgt Ion:113 Resp: 49246
Ion Ratio Lower Upper
113 100
111 102.1 83.4 125.0
192 17.0 14.1 21.1









#50
Toluene-d8
Concen: 50.286 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX043357.D
Acq: 11 Oct 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :
VX1011MBL01


Tgt Ion: 98 Resp: 177332
Ion Ratio Lower Upper
98 100
100 65.8 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 48.932 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX043357.D
Acq: 11 Oct 2024 10:55

Tgt Ion: 95 Resp: 62689
Ion Ratio Lower Upper
95 100
174 72.2 0.0 145.0
176 70.7 0.0 143.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX043357.D

Acq: 11 Oct 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :

VX1011MBL01





Tgt Ion:117 Resp: 127945

Ion Ratio Lower Upper

117 100

82 56.8 45.0 67.6

119 32.0 25.3 37.9



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX043357.D

Acq: 11 Oct 2024 10:55

Tgt Ion:152 Resp: 53943

Ion Ratio Lower Upper

152 100

115 63.6 44.1 132.3

150 157.9 0.0 347.4

