

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101218\
 Data File : VX005314.D
 Acq On : 13 Oct 2018 01:24
 Operator : JC/MD
 Sample : J5393-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-DUP05-20181008DL

Quant Time: Oct 13 05:15:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	288576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	389237	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	349895	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	171202	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	158450	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
35) Dibromofluoromethane	5.50	113	126442	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
50) Toluene-d8	8.72	98	467414	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
62) 4-Bromofluorobenzene	11.14	95	151339	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
Target Compounds						
44) Trichloroethene	7.21	130	114027	39.36	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101218\
Data File : VX005314.D
Acq On : 13 Oct 2018 01:24
Operator : JC/MD
Sample : J5393-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-DUP05-20181008DL

Quant Time: Oct 13 05:15:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

