

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101218\
 Data File : VX005322.D
 Acq On : 13 Oct 2018 04:57
 Operator : JC/MD
 Sample : J5392-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-DRUM

Quant Time: Oct 13 05:38:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	384760	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	350567	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167066	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	154260	52.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.12%	
35) Dibromofluoromethane	5.51	113	122434	48.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.08%	
50) Toluene-d8	8.71	98	461684	52.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.54%	
62) 4-Bromofluorobenzene	11.14	95	150567	45.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	20535	12.10	ug/l	93
18) Methyl Acetate	2.78	43	9098	1.88	ug/l	96
20) Methylene Chloride	2.86	84	245502	100.61	ug/l	96
25) 2-Butanone	4.70	43	3864	1.56	ug/l	94
43) Isopropyl Acetate	6.46	43	148532	23.27	ug/l	96
68) m/p-Xylenes	10.36	106	5806	1.23	ug/l	94
84) 1,2,4-Trimethylbenzene	11.81	105	17927	2.05	ug/l	97
95) Naphthalene	13.83	128	289576	25.13	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101218\
Data File : VX005322.D
Acq On : 13 Oct 2018 04:57
Operator : JC/MD
Sample : J5392-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-DRUM

Quant Time: Oct 13 05:38:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

