

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018906.D
 Acq On : 12 Oct 2020 16:22
 Operator : JC/SP
 Sample : L4303-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-MW03S-20201004

Quant Time: Oct 13 13:57:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	247328	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	389397	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	390567	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	204339	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	193143	52.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.90%	
35) Dibromofluoromethane	5.47	113	145874	52.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.14%	
50) Toluene-d8	8.70	98	490500	50.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.92%	
62) 4-Bromofluorobenzene	11.12	95	182256	47.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.02%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018906.D
Acq On : 12 Oct 2020 16:22
Operator : JC/SP
Sample : L4303-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-MW03S-20201004

Quant Time: Oct 13 13:57:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

