

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018907.D
 Acq On : 12 Oct 2020 16:44
 Operator : JC/SP
 Sample : L4303-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-05-20201002

Quant Time: Oct 13 04:47:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	242673	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	399005	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	387976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	200953	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	194050	54.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.44%	
35) Dibromofluoromethane	5.47	113	144978	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
50) Toluene-d8	8.70	98	490418	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
62) 4-Bromofluorobenzene	11.12	95	185373	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
Target Compounds						
3) Chloromethane	1.31	50	2521	1.108	ug/l	96
20) Methylene Chloride	2.84	84	1303	1.009	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018907.D
Acq On : 12 Oct 2020 16:44
Operator : JC/SP
Sample : L4303-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-05-20201002

Quant Time: Oct 13 04:47:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

