

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018910.D
 Acq On : 12 Oct 2020 17:53
 Operator : JC/SP
 Sample : PB132255ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132255ZHE#02

Quant Time: Oct 13 04:56:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	256672	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	426715	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	464135	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	217577	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	206339	54.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.02%	
35) Dibromofluoromethane	5.47	113	154302	50.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.48%	
50) Toluene-d8	8.70	98	565489	53.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.22%	
62) 4-Bromofluorobenzene	11.12	95	210877	50.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.32%	
Target Compounds						
18) Methyl Acetate	2.75	43	3823	1.215	ug/l	92
20) Methylene Chloride	2.84	84	5824	2.639	ug/l	88
43) Isopropyl Acetate	6.41	43	77519	10.766	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018910.D
Acq On : 12 Oct 2020 17:53
Operator : JC/SP
Sample : PB132255ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132255ZHE#02

Quant Time: Oct 13 04:56:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

