

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018913.D
 Acq On : 12 Oct 2020 19:01
 Operator : JC/SP
 Sample : PB132255ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132255ZHE#05

Quant Time: Oct 13 05:01:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	247297	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	380939	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	411854	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	198217	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	190870	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
35) Dibromofluoromethane	5.46	113	146279	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
50) Toluene-d8	8.70	98	501543	53.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.52%	
62) 4-Bromofluorobenzene	11.12	95	183421	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
Target Compounds						
20) Methylene Chloride	2.83	84	5800	2.711	ug/l	95
43) Isopropyl Acetate	6.42	43	62769	9.765	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018913.D
Acq On : 12 Oct 2020 19:01
Operator : JC/SP
Sample : PB132255ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132255ZHE#05

Quant Time: Oct 13 05:01:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

