

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018914.D
 Acq On : 12 Oct 2020 19:23
 Operator : JC/SP
 Sample : PB132255ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132255ZHE#06

Quant Time: Oct 13 05:02:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	252037	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	402980	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	414802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	207889	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	185724	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
35) Dibromofluoromethane	5.47	113	144740	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
50) Toluene-d8	8.70	98	522275	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
62) 4-Bromofluorobenzene	11.12	95	193959	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
Target Compounds						
20) Methylene Chloride	2.84	84	5434	2.533	ug/l	91
43) Isopropyl Acetate	6.42	43	61896	9.103	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018914.D
Acq On : 12 Oct 2020 19:23
Operator : JC/SP
Sample : PB132255ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132255ZHE#06

Quant Time: Oct 13 05:02:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

