

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018915.D
 Acq On : 12 Oct 2020 19:46
 Operator : JC/SP
 Sample : PB132255ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132255ZHE#07

Quant Time: Oct 13 05:04:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	250584	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	392527	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	420485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206805	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	186486	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
35) Dibromofluoromethane	5.46	113	141850	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
50) Toluene-d8	8.70	98	518161	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	
62) 4-Bromofluorobenzene	11.12	95	190626	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
Target Compounds						
20) Methylene Chloride	2.83	84	4397	2.155	ug/l	89
43) Isopropyl Acetate	6.42	43	63557	9.596	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018915.D
Acq On : 12 Oct 2020 19:46
Operator : JC/SP
Sample : PB132255ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132255ZHE#07

Quant Time: Oct 13 05:04:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

