

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018916.D
 Acq On : 12 Oct 2020 20:09
 Operator : JC/SP
 Sample : PB132255ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132255ZHE#08

Quant Time: Oct 13 05:06:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	241931	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	383573	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	403576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	198548	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	183448	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
35) Dibromofluoromethane	5.46	113	138906	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
50) Toluene-d8	8.70	98	509830	53.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.54%	
62) 4-Bromofluorobenzene	11.12	95	187846	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
Target Compounds						
18) Methyl Acetate	2.75	43	3059	1.031	ug/l #	89
20) Methylene Chloride	2.84	84	3598	1.903	ug/l	92
43) Isopropyl Acetate	6.42	43	67340	10.404	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018916.D
Acq On : 12 Oct 2020 20:09
Operator : JC/SP
Sample : PB132255ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132255ZHE#08

Quant Time: Oct 13 05:06:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

