

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018917.D
 Acq On : 12 Oct 2020 20:32
 Operator : JC/SP
 Sample : L4276-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 13 05:09:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	252899	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	392518	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	405473	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	196705	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	187039	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
35) Dibromofluoromethane	5.46	113	144158	51.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.06%	
50) Toluene-d8	8.70	98	493397	50.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.70%	
62) 4-Bromofluorobenzene	11.12	95	180591	46.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.40%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018917.D
Acq On : 12 Oct 2020 20:32
Operator : JC/SP
Sample : L4276-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 13 05:09:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

