

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018918.D
 Acq On : 12 Oct 2020 20:54
 Operator : JC/SP
 Sample : L4276-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Oct 13 05:11:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	241195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	389832	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	383511	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195042	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	189755	53.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.68%	
35) Dibromofluoromethane	5.46	113	145384	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
50) Toluene-d8	8.70	98	477783	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.12	95	174154	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
Target Compounds						
16) Acetone	2.42	43	16467	7.228	ug/l	94
20) Methylene Chloride	2.83	84	2170	1.350	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018918.D
Acq On : 12 Oct 2020 20:54
Operator : JC/SP
Sample : L4276-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Oct 13 05:11:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

