

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018919.D
 Acq On : 12 Oct 2020 21:17
 Operator : JC/SP
 Sample : L4276-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Oct 13 05:13:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	248568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	384070	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	390672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	197706	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	194834	53.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.28%	
35) Dibromofluoromethane	5.46	113	144296	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
50) Toluene-d8	8.70	98	493959	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.12	95	178468	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
Target Compounds						
16) Acetone	2.42	43	16705	7.032	ug/l #	88
20) Methylene Chloride	2.84	84	2083	1.292	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018919.D
Acq On : 12 Oct 2020 21:17
Operator : JC/SP
Sample : L4276-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Oct 13 05:13:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

