

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101220\
 Data File : VX018920.D
 Acq On : 12 Oct 2020 21:40
 Operator : JC/SP
 Sample : L4299-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 86-96-COMP

Quant Time: Oct 13 05:17:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	248859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	388822	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	391605	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	210012	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	181750	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
35) Dibromofluoromethane	5.47	113	143191	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
50) Toluene-d8	8.70	98	485147	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	11.12	95	188834	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
Target Compounds						
11) Tert butyl alcohol	3.01	59	62820	108.669	ug/l	99
17) Carbon Disulfide	2.56	76	2610	0.387	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101220\
Data File : VX018920.D
Acq On : 12 Oct 2020 21:40
Operator : JC/SP
Sample : L4299-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
86-96-COMP

Quant Time: Oct 13 05:17:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

