

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
 Data File : VX024760.D
 Acq On : 12 Oct 2021 14:00
 Operator : JC/MD
 Sample : VX1012MBS01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 12 14:23:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	147560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231223	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109463	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109455	52.63	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.26%			
35) Dibromofluoromethane	5.391	113	80692	52.53	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.06%			
50) Toluene-d8	8.653	98	288384	51.01	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.02%			
62) 4-Bromofluorobenzene	11.085	95	111424	50.13	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.26%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	36821	24.80	ug/l	99
3) Chloromethane	1.295	50	26538	18.96	ug/l	99
4) Vinyl Chloride	1.374	62	26723	18.25	ug/l	97
5) Bromomethane	1.605	94	20180	20.48	ug/l	100
6) Chloroethane	1.685	64	16552	17.15	ug/l	97
7) Trichlorofluoromethane	1.886	101	44683	17.30	ug/l	96
8) Diethyl Ether	2.136	74	14501	16.49	ug/l	92
9) 1,1,2-Trichlorotrifluo...	2.331	101	24466	16.79	ug/l	96
10) Methyl Iodide	2.453	142	26769	13.13	ug/l	92
11) Tert butyl alcohol	2.971	59	33569	60.00	ug/l	98
12) 1,1-Dichloroethene	2.319	96	23191	17.14	ug/l	93
13) Acrolein	2.239	56	10371	98.55	ug/l	94
14) Allyl chloride	2.666	41	40130	16.23	ug/l	90
15) Acrylonitrile	3.069	53	76788	81.27	ug/l	99
16) Acetone	2.386	43	93215	100.15	ug/l	88
17) Carbon Disulfide	2.514	76	52670	15.26	ug/l	98
18) Methyl Acetate	2.709	43	51979	16.69	ug/l	91
19) Methyl tert-butyl Ether	3.123	73	82456	16.18	ug/l	99
20) Methylene Chloride	2.794	84	27466	16.89	ug/l	96
21) trans-1,2-Dichloroethene	3.099	96	25490	16.60	ug/l	99
22) Diisopropyl ether	3.770	45	82068	16.24	ug/l	97
23) Vinyl Acetate	3.733	43	335256	81.94	ug/l	94
24) 1,1-Dichloroethane	3.611	63	46609	16.61	ug/l	95
25) 2-Butanone	4.568	43	131106	91.54	ug/l	91
26) 2,2-Dichloropropane	4.489	77	40023	15.86	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	28531	15.89	ug/l	93
28) Bromochloromethane	4.904	49	22106	17.36	ug/l	99
29) Tetrahydrofuran	5.026	42	73802	80.87	ug/l	90
30) Chloroform	5.105	83	50263	16.22	ug/l	95
31) Cyclohexane	5.477	56	40322	16.19	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	45388	16.50	ug/l	99
36) 1,1-Dichloropropene	5.702	75	35482	16.01	ug/l	99
37) Ethyl Acetate	4.727	43	42291	15.28	ug/l	96
38) Carbon Tetrachloride	5.684	117	38829	16.09	ug/l	98
39) Methylcyclohexane	7.385	83	41257	15.90	ug/l	98
40) Benzene	6.044	78	104317	16.46	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
 Data File : VX024760.D
 Acq On : 12 Oct 2021 14:00
 Operator : JC/MD
 Sample : VX1012MBS01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 12 14:23:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	23318	16.32	ug/l	92
42) 1,2-Dichloroethane	6.092	62	44174	16.59	ug/l	93
43) Isopropyl Acetate	6.355	43	66028	15.69	ug/l	94
44) Trichloroethene	7.129	130	29306	16.80	ug/l	93
45) 1,2-Dichloropropane	7.434	63	26525	16.22	ug/l	94
46) Dibromomethane	7.586	93	19329	15.76	ug/l	98
47) Bromodichloromethane	7.830	83	35361	15.73	ug/l	99
48) Methyl methacrylate	7.696	41	35635	17.10	ug/l	88
49) 1,4-Dioxane	7.665	88	13928	291.88	ug/l #	93
51) 4-Methyl-2-Pentanone	8.580	43	251878	90.68	ug/l	91
52) Toluene	8.720	92	68354	16.69	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	38307	15.04	ug/l	98
54) cis-1,3-Dichloropropene	8.373	75	41605	16.06	ug/l #	88
55) 1,1,2-Trichloroethane	9.159	97	27777	16.41	ug/l	96
56) Ethyl methacrylate	9.122	69	37028	13.74	ug/l #	83
57) 1,3-Dichloropropane	9.311	76	46628	16.32	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	126309	94.96	ug/l	98
59) 2-Hexanone	9.433	43	199793	91.34	ug/l	86
60) Dibromochloromethane	9.525	129	26642	14.73	ug/l	98
61) 1,2-Dibromoethane	9.610	107	30055	16.17	ug/l	98
64) Tetrachloroethene	9.275	164	27811	18.63	ug/l	94
65) Chlorobenzene	10.086	112	73518	16.36	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	26784	15.99	ug/l	100
67) Ethyl Benzene	10.195	91	129694	16.30	ug/l	96
68) m/p-Xylenes	10.305	106	99234	32.53	ug/l	98
69) o-Xylene	10.646	106	47750	15.83	ug/l	98
70) Styrene	10.659	104	76165	15.10	ug/l	96
71) Bromoform	10.805	173	17566	13.80	ug/l #	97
73) Isopropylbenzene	10.964	105	131095	16.94	ug/l	100
74) N-amyl acetate	10.848	43	54805	15.40	ug/l	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	43674	15.96	ug/l	97
76) 1,2,3-Trichloropropane	11.244	75	49561	19.94	ug/l	91
77) Bromobenzene	11.201	156	32228	16.18	ug/l	97
78) n-propylbenzene	11.305	91	144437	15.97	ug/l	99
79) 2-Chlorotoluene	11.366	91	89947	16.32	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	113727	17.20	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.024	75	12351	15.10	ug/l	87
82) 4-Chlorotoluene	11.457	91	106099	16.36	ug/l	99
83) tert-Butylbenzene	11.719	119	104964	15.84	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	113970	17.34	ug/l	98
85) sec-Butylbenzene	11.896	105	124577	15.28	ug/l	99
86) p-Isopropyltoluene	12.012	119	112459	16.17	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	57937	15.28	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	58819	15.01	ug/l	97
89) n-Butylbenzene	12.335	91	96401	15.66	ug/l	98
90) Hexachloroethane	12.542	117	16340	14.29	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	58619	15.64	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.945	75	9781	14.80	ug/l	90
93) 1,2,4-Trichlorobenzene	13.591	180	36036	15.10	ug/l	99
94) Hexachlorobutadiene	13.731	225	16654	15.25	ug/l	100
95) Naphthalene	13.780	128	119182	15.60	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	36363	14.87	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
Data File : VX024760.D
Acq On : 12 Oct 2021 14:00
Operator : JC/MD
Sample : VX1012MBS01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Oct 12 14:23:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 12 14:23:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 147560
Ion Ratio Lower Upper
168 100
99 58.2 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 24.80 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 85 Resp: 36821
Ion Ratio Lower Upper
85 100
87 33.5 16.4 49.2









#3
Chloromethane
Concen: 18.96 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 26538
Ion Ratio Lower Upper
50 100
52 32.4 26.4 39.6



#4
Vinyl Chloride
Concen: 18.25 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 62 Resp: 26723
Ion Ratio Lower Upper
62 100
64 33.5 25.7 38.5









#5
Bromomethane
Concen: 20.48 ug/l
RT: 1.605 min Scan# 85
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 20180
Ion Ratio Lower Upper
94 100
96 95.9 76.4 114.6



#6
Chloroethane
Concen: 17.15 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 64 Resp: 16552
Ion Ratio Lower Upper
64 100
66 33.3 25.4 38.0









#7
Trichlorofluoromethane
Concen: 17.30 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 101 Resp: 44683
Ion Ratio Lower Upper
101 100
103 63.4 53.2 79.8



#8
Diethyl Ether
Concen: 16.49 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 74 Resp: 14501
Ion Ratio Lower Upper
74 100
45 96.8 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 16.79 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	101	Resp:	24466
Ion	Ratio	Lower	Upper
101	100		
85	46.4	34.9	52.3
151	80.5	61.4	92.2



#10
Methyl Iodide
Concen: 13.13 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	142	Resp:	26769
Ion	Ratio	Lower	Upper
142	100		
127	48.4	33.3	49.9
141	14.1	11.4	17.0









#11
Tert butyl alcohol
Concen: 60.00 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 59 Resp: 33569
Ion Ratio Lower Upper
59 100
57 10.4 7.8 11.8



#12
1,1-Dichloroethene
Concen: 17.14 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 96 Resp: 23191
Ion Ratio Lower Upper
96 100
61 167.7 124.6 187.0
98 62.8 50.6 76.0









#13
Acrolein
Concen: 98.55 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 10371
Ion Ratio Lower Upper
56 100
55 75.5 56.4 84.6



#14
Allyl chloride
Concen: 16.23 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 41 Resp: 40130
Ion Ratio Lower Upper
41 100
39 71.6 51.6 77.4
76 30.9 30.5 45.7









#15
Acrylonitrile
Concen: 81.27 ug/l
RT: 3.069 min Scan# 325
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	53	Resp:	76788
Ion	Ratio	Lower	Upper
53	100		
52	81.2	65.3	97.9
51	35.9	28.2	42.2



#16
Acetone
Concen: 100.15 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	43	Resp:	93215
Ion	Ratio	Lower	Upper
43	100		
58	28.0	27.8	41.8









#17
Carbon Disulfide
Concen: 15.26 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 52670
Ion Ratio Lower Upper
76 100
78 9.6 7.2 10.8



#18
Methyl Acetate
Concen: 16.69 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 43 Resp: 51979
Ion Ratio Lower Upper
43 100
74 22.7 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 16.18 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 82456
Ion Ratio Lower Upper
73 100
57 21.7 17.1 25.7



#20
Methylene Chloride
Concen: 16.89 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 84 Resp: 27466
Ion Ratio Lower Upper
84 100
49 116.8 88.6 132.8
51 36.5 27.0 40.6
86 62.5 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 16.60 ug/l
RT: 3.099 min Scan# 330
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 96 Resp: 25490




Ion	Ratio	Lower	Upper
96	100		
61	134.9	107.7	161.5
98	60.8	50.4	75.6




#22
Diisopropyl ether
Concen: 16.24 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 45 Resp: 82068

Ion	Ratio	Lower	Upper
45	100		
43	73.6	58.0	87.0
87	27.4	25.4	38.0
59	11.5	10.2	15.2









#23
 Vinyl Acetate
 Concen: 81.94 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. 0.000 min
 Lab File: VX024760.D
 Acq: 12 Oct 2021 14:00




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 335256
 Ion Ratio Lower Upper
 43 100
 86 10.0 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 16.61 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX024760.D
 Acq: 12 Oct 2021 14:00

Tgt Ion: 63 Resp: 46609
 Ion Ratio Lower Upper
 63 100
 98 5.2 3.8 11.2
 100 3.7 2.5 7.5









#25
2-Butanone
Concen: 91.54 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 43 Resp: 131106
Ion Ratio Lower Upper
43 100
72 24.7 23.7 35.5




#26
2,2-Dichloropropane
Concen: 15.86 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 77 Resp: 40023
Ion Ratio Lower Upper
77 100
97 21.8 11.7 35.1



Abundance Scan 559 (4.495 min): VX024453.D\data.ms (-547





#27
cis-1,2-Dichloroethene
Concen: 15.89 ug/l
RT: 4.495 min Scan# 559
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00


Tgt Ion:	96	Resp:	28531
Ion	Ratio	Lower	Upper
96	100		
61	155.1	0.0	285.6
98	65.4	0.0	129.6

Instrument :
MSVOA_X
ClientSampled :


Abundance Scan 559 (4.495 min): VX024760.D\data.ms



Abundance Scan 559 (4.495 min): VX024760.D\data.ms (-510





Abundance Scan 627 (4.910 min): VX024453.D\data.ms (-614




#28
Bromochloromethane
Concen: 17.36 ug/l
RT: 4.904 min Scan# 626
Delta R.T. -0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00


Tgt Ion:	49	Resp:	22106
Ion	Ratio	Lower	Upper
49	100		
129	2.5	0.0	4.0
130	84.4	68.2	102.2

Abundance Scan 626 (4.904 min): VX024760.D\data.ms



Abundance Scan 626 (4.904 min): VX024760.D\data.ms (-578









#29
Tetrahydrofuran
Concen: 80.87 ug/l
RT: 5.026 min Scan# 646
Delta R.T. 0.007 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	42	Resp:	73802
Ion	Ratio	Lower	Upper
42	100		
72	44.6	41.7	62.5
71	41.2	38.4	57.6



#30
Chloroform
Concen: 16.22 ug/l
RT: 5.105 min Scan# 659
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	83	Resp:	50263
Ion	Ratio	Lower	Upper
83	100		
85	61.4	52.0	78.0









#31
Cyclohexane
Concen: 16.19 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 40322
Ion Ratio Lower Upper
56 100
69 28.7 25.3 37.9
84 86.4 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 16.50 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 97 Resp: 45388
Ion Ratio Lower Upper
97 100
99 64.9 51.5 77.3
61 43.6 34.2 51.4









#33
1,2-Dichloroethane-d4
Concen: 52.63 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 65 Resp: 109455
Ion Ratio Lower Upper
65 100
67 49.8 0.0 105.4



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 114 Resp: 231223
Ion Ratio Lower Upper
114 100
63 20.7 0.0 40.8
88 16.4 0.0 32.8









#35
Dibromofluoromethane
Concen: 52.53 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:113 Resp: 80692
Ion Ratio Lower Upper
113 100
111 102.9 81.9 122.9
192 19.9 16.7 25.1



#36
1,1-Dichloropropene
Concen: 16.01 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 75 Resp: 35482
Ion Ratio Lower Upper
75 100
110 36.0 18.1 54.4
77 30.7 25.3 37.9









#37
Ethyl Acetate
Concen: 15.28 ug/l
RT: 4.727 min Scan# 597
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:	43	Resp:	42291
Ion	Ratio	Lower	Upper
43	100		
61	13.9	12.8	19.2
70	11.6	10.5	15.7




#38
Carbon Tetrachloride
Concen: 16.09 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	117	Resp:	38829
Ion	Ratio	Lower	Upper
117	100		
119	93.3	76.2	114.2
121	30.8	24.8	37.2



Abundance Scan 1033 (7.385 min): VX024453.D\data.ms (-16)





#39
Methylcyclohexane
Concen: 15.90 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00


Tgt Ion:	83	Resp:	41257
Ion	Ratio	Lower	Upper
83	100		
55	76.4	58.9	88.3
98	47.5	38.1	57.1

Instrument :
MSVOA_X
ClientSampled :


Abundance Scan 1033 (7.385 min): VX024760.D\data.ms



Abundance Scan 1033 (7.385 min): VX024760.D\data.ms (-96)





Abundance Scan 813 (6.044 min): VX024453.D\data.ms (-80)




#40
Benzene
Concen: 16.46 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	78	Resp:	104317
Ion	Ratio	Lower	Upper
78	100		
77	23.5	18.9	28.3


Abundance Scan 813 (6.044 min): VX024760.D\data.ms



Abundance Scan 813 (6.044 min): VX024760.D\data.ms (-76)




Abundance Scan 630 (4.928 min): VX024453.D\data.ms (-618





#41
Methacrylonitrile
Concen: 16.32 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00


Tgt Ion:	41	Resp:	23318
Ion	Ratio	Lower	Upper
41	100		
39	56.5	42.5	63.7
67	70.8	65.0	97.6
52	29.4	25.3	37.9



Abundance Scan 630 (4.928 min): VX024760.D\data.ms (-58




Abundance Scan 821 (6.092 min): VX024453.D\data.ms (-80





#42
1,2-Dichloroethane
Concen: 16.59 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00


Tgt Ion:	62	Resp:	44174
Ion	Ratio	Lower	Upper
62	100		
98	7.5	0.0	20.2



Abundance Scan 821 (6.092 min): VX024760.D\data.ms (-77







Instrument :
MSVOA_X
ClientSampled :



#43
Isopropyl Acetate
Concen: 15.69 ug/l
RT: 6.355 min Scan# 864
Delta R.T. 0.007 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	43	Resp:	66028
Ion	Ratio	Lower	Upper
43	100		
61	19.2	18.2	27.4
87	13.6	12.5	18.7



#44
Trichloroethene
Concen: 16.80 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	130	Resp:	29306
Ion	Ratio	Lower	Upper
130	100		
95	92.3	0.0	198.0









#45
1,2-Dichloropropane
Concen: 16.22 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 63 Resp: 26525
Ion Ratio Lower Upper
63 100
65 29.0 25.7 38.5



#46
Dibromomethane
Concen: 15.76 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 93 Resp: 19329
Ion Ratio Lower Upper
93 100
95 84.3 66.8 100.2
174 100.4 83.1 124.7





#47

Bromodichloromethane

Concen: 15.73 ug/l

RT: 7.830 min Scan# 1106





Delta R.T. -0.000 min

Lab File: VX024760.D

Acq: 12 Oct 2021 14:00

Tgt Ion:	83	Resp:	35361
Ion	Ratio	Lower	Upper
83	100		
85	63.5	51.4	77.0
127	8.8	6.8	10.2

Instrument :
MSVOA_X
ClientSampled :



#48

Methyl methacrylate

Concen: 17.10 ug/l




RT: 7.696 min Scan# 1084

Delta R.T. -0.000 min


Lab File: VX024760.D

Acq: 12 Oct 2021 14:00

Tgt Ion:	41	Resp:	35635
Ion	Ratio	Lower	Upper
41	100		
69	79.4	76.0	114.0
39	59.6	44.6	67.0





Abundance Scan 1079 (7.665 min): VX024453.D\data.ms (-10




#49
1,4-Dioxane
Concen: 291.88 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	88	Resp:	13928
Ion	Ratio	Lower	Upper
88	100		
43	37.4	23.2	34.8#
58	71.6	55.0	82.4


Instrument :
MSVOA_X
ClientSampled :



Abundance Scan 1079 (7.665 min): VX024760.D\data.ms (-10





Abundance Scan 1241 (8.653 min): VX024453.D\data.ms (-12





#50
Toluene-d8
Concen: 51.01 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	98	Resp:	288384
Ion	Ratio	Lower	Upper
98	100		
100	65.0	51.9	77.9



Abundance Scan 1241 (8.653 min): VX024760.D\data.ms (-12









#51
4-Methyl-2-Pentanone
Concen: 90.68 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 251878
Ion Ratio Lower Upper
43 100
58 38.2 35.4 53.0



#52
Toluene
Concen: 16.69 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 92 Resp: 68354
Ion Ratio Lower Upper
92 100
91 173.2 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 15.04 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 38307
Ion Ratio Lower Upper
75 100
77 32.1 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 16.06 ug/l
RT: 8.373 min Scan# 1195
Delta R.T. 0.001 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 75 Resp: 41605
Ion Ratio Lower Upper
75 100
77 34.8 25.7 38.5
39 54.8 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 16.41 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	97	Resp:	27777
Ion	Ratio	Lower	Upper
97	100		
83	81.1	69.3	103.9
85	58.4	44.3	66.5
99	63.6	50.3	75.5



#56
Ethyl methacrylate
Concen: 13.74 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	69	Resp:	37028
Ion	Ratio	Lower	Upper
69	100		
41	69.2	46.8	70.2
39	42.3	24.7	37.1#









#57
1,3-Dichloropropane
Concen: 16.32 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 46628
Ion Ratio Lower Upper
76 100
78 32.3 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 94.96 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 63 Resp: 126309
Ion Ratio Lower Upper
63 100
106 28.2 23.5 35.3









#59
2-Hexanone
Concen: 91.34 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 199793
Ion Ratio Lower Upper
43 100
58 51.3 30.9 92.8



#60
Dibromochloromethane
Concen: 14.73 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 129 Resp: 26642
Ion Ratio Lower Upper
129 100
127 78.3 38.4 115.2









#61
1,2-Dibromoethane
Concen: 16.17 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 30055
Ion Ratio Lower Upper
107 100
109 92.7 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 50.13 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 95 Resp: 111424
Ion Ratio Lower Upper
95 100
174 80.4 0.0 154.6
176 78.6 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	117	Resp:	215724
Ion	Ratio	Lower	Upper
117	100		
82	58.8	47.4	71.2
119	32.8	26.6	39.8



#64
Tetrachloroethene
Concen: 18.63 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	164	Resp:	27811
Ion	Ratio	Lower	Upper
164	100		
166	126.4	103.4	155.2
129	99.2	70.8	106.2
131	95.9	72.5	108.7









#65
Chlorobenzene
Concen: 16.36 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 73518
Ion Ratio Lower Upper
112 100
114 31.0 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 15.99 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:131 Resp: 26784
Ion Ratio Lower Upper
131 100
133 95.2 47.7 143.1
119 64.9 32.1 96.5









#67
Ethyl Benzene
Concen: 16.30 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 129694
Ion Ratio Lower Upper
91 100
106 33.2 24.6 37.0



#68
m/p-Xylenes
Concen: 32.53 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:106 Resp: 99234
Ion Ratio Lower Upper
106 100
91 207.3 162.8 244.2









#69
o-Xylene
Concen: 15.83 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 47750
Ion Ratio Lower Upper
106 100
91 218.6 108.1 324.2



#70
Styrene
Concen: 15.10 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:104 Resp: 76165
Ion Ratio Lower Upper
104 100
78 53.4 40.0 60.0
103 56.9 43.4 65.0









#71
Bromoform
Concen: 13.80 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion	Ratio	Lower	Upper
173	100		
175	50.0	23.9	71.8
254	0.0	0.1	0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion	Ratio	Lower	Upper
152	100		
115	68.7	41.1	123.3
150	162.0	0.0	349.0









#73
Isopropylbenzene
Concen: 16.94 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.001 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:105 Resp: 131095
Ion Ratio Lower Upper
105 100
120 26.3 13.2 39.6



#74
N-ethyl acetate
Concen: 15.40 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 43 Resp: 54805
Ion Ratio Lower Upper
43 100
70 41.3 39.0 58.6
55 24.7 21.5 32.3
61 23.2 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 15.96 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	83	Resp:	43674
Ion	Ratio	Lower	Upper
83	100		
131	10.5	5.1	15.2
85	66.3	31.8	95.4



#76
1,2,3-Trichloropropane
Concen: 19.94 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	75	Resp:	49561
Ion	Ratio	Lower	Upper
75	100		
77	32.7	19.1	57.5









#77
Bromobenzene
Concen: 16.18 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:156 Resp: 32228
Ion Ratio Lower Upper
156 100
77 161.4 82.8 248.6
158 99.9 49.1 147.4



#78
n-propylbenzene
Concen: 15.97 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion: 91 Resp: 144437
Ion Ratio Lower Upper
91 100
120 23.5 11.5 34.5









#79
2-Chlorotoluene
Concen: 16.32 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 89947
Ion Ratio Lower Upper
91 100
126 33.7 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 17.20 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:105 Resp: 113727
Ion Ratio Lower Upper
105 100
120 45.9 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 15.10 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	75	Resp:	12351
Ion	Ratio	Lower	Upper
75	100		
53	110.0	74.5	111.7
89	44.6	37.9	56.9



#82
4-Chlorotoluene
Concen: 16.36 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	91	Resp:	106099
Ion	Ratio	Lower	Upper
91	100		
126	28.9	14.9	44.5









#83
tert-Butylbenzene
Concen: 15.84 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	119	Resp:	104964
Ion	Ratio	Lower	Upper
119	100		
91	61.8	27.9	83.7
134	23.5	11.3	33.9



#84
1,2,4-Trimethylbenzene
Concen: 17.34 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	105	Resp:	113970
Ion	Ratio	Lower	Upper
105	100		
120	43.8	22.4	67.3









#85
 sec-Butylbenzene
 Concen: 15.28 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX024760.D
 Acq: 12 Oct 2021 14:00




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 124577
 Ion Ratio Lower Upper
 105 100
 134 20.1 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 16.17 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX024760.D
 Acq: 12 Oct 2021 14:00

Tgt Ion:119 Resp: 112459
 Ion Ratio Lower Upper
 119 100
 134 26.1 12.9 38.7
 91 26.3 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 15.28 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 57937
Ion Ratio Lower Upper
146 100
111 41.1 20.2 60.5
148 65.5 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 15.01 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:146 Resp: 58819
Ion Ratio Lower Upper
146 100
111 42.9 19.6 58.8
148 63.6 31.6 95.0









#89
n-Butylbenzene
Concen: 15.66 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	91	Resp:	96401
Ion	Ratio	Lower	Upper
91	100		
92	52.9	27.4	82.0
134	26.6	13.2	39.5



#90
Hexachloroethane
Concen: 14.29 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	117	Resp:	16340
Ion	Ratio	Lower	Upper
117	100		
201	75.7	40.4	121.2









#91
1,2-Dichlorobenzene
Concen: 15.64 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	146	Resp:	58619
Ion	Ratio	Lower	Upper
146	100		
111	41.7	20.8	62.2
148	63.4	31.9	95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 14.80 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	75	Resp:	9781
Ion	Ratio	Lower	Upper
75	100		
155	83.9	45.3	135.9
157	102.8	57.7	173.1









#93
1,2,4-Trichlorobenzene
Concen: 15.10 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.2	47.7	143.1
145	31.2	15.0	45.0



#94
Hexachlorobutadiene
Concen: 15.25 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.1	31.1	93.2
227	63.4	31.6	95.0









#95
Naphthalene
Concen: 15.60 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:	128	Resp:	119182
Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.3	15.5
129	10.7	8.8	13.2



#96
1,2,3-Trichlorobenzene
Concen: 14.87 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX024760.D
Acq: 12 Oct 2021 14:00

Tgt Ion:	180	Resp:	36363
Ion	Ratio	Lower	Upper
180	100		
182	96.6	47.6	142.9
145	34.6	16.1	48.3

