

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
 Data File : VX024766.D
 Acq On : 12 Oct 2021 16:25
 Operator : JC/MD
 Sample : M4170-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P287-DITCH

Quant Time: Oct 12 16:48:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	225005	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98702	54.391	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.780%	
35) Dibromofluoromethane	5.397	113	77619	51.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.860%	
50) Toluene-d8	8.653	98	276385	50.238	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.480%	
62) 4-Bromofluorobenzene	11.085	95	108457	50.140	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.280%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
Data File : VX024766.D
Acq On : 12 Oct 2021 16:25
Operator : JC/MD
Sample : M4170-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P287-DITCH

Quant Time: Oct 12 16:48:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

