

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101221\
 Data File : VX024756.D
 Acq On : 12 Oct 2021 10:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 01:45:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	58.021	-16.0	112	0.00
3 P	Chloromethane	50.000	48.303	3.4	96	0.00
4 C	Vinyl Chloride	50.000	49.385	1.2#	97	0.00
5 T	Bromomethane	50.000	57.949	-15.9	113	0.00
6 T	Chloroethane	50.000	46.902	6.2	97	0.00
7 T	Trichlorofluoromethane	50.000	49.166	1.7	98	0.00
8 T	Diethyl Ether	50.000	47.142	5.7	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.188	-0.4	101	0.00
10 T	Methyl Iodide	50.000	43.398	13.2	86	0.00
11 T	Tert butyl alcohol	250.000	180.390	27.8#	72	0.00
12 CM	1,1-Dichloroethene	50.000	49.178	1.6#	100	0.00
13 T	Acrolein	250.000	231.315	7.5	91	0.00
14 T	Allyl chloride	50.000	48.302	3.4	96	0.00
15 T	Acrylonitrile	250.000	224.659	10.1	91	0.00
16 T	Acetone	250.000	346.701	-38.7#	131	0.00
17 T	Carbon Disulfide	50.000	45.865	8.3	95	0.00
18 T	Methyl Acetate	50.000	47.141	5.7	98	0.00
19 T	Methyl tert-butyl Ether	50.000	47.470	5.1	96	0.00
20 T	Methylene Chloride	50.000	49.501	1.0	101	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.747	2.5	100	0.00
22 T	Diisopropyl ether	50.000	48.251	3.5	97	0.00
23 T	Vinyl Acetate	250.000	247.082	1.2	96	0.00
24 P	1,1-Dichloroethane	50.000	48.752	2.5	98	0.00
25 T	2-Butanone	250.000	270.423	-8.2	107	0.00
26 T	2,2-Dichloropropane	50.000	50.552	-1.1	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.091	3.8	98	0.00
28 T	Bromochloromethane	50.000	43.664	12.7	97	-0.01
29 T	Tetrahydrofuran	250.000	222.529	11.0	87	0.00
30 C	Chloroform	50.000	48.504	3.0#	98	0.00
31 T	Cyclohexane	50.000	48.329	3.3	96	0.00
32 T	1,1,1-Trichloroethane	50.000	50.294	-0.6	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.651	0.7	101	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	52.325	-4.7	103	0.00
36 T	1,1-Dichloropropene	50.000	50.361	-0.7	99	0.00
37 T	Ethyl Acetate	50.000	46.178	7.6	92	0.00
38 T	Carbon Tetrachloride	50.000	52.056	-4.1	98	-0.01
39 T	Methylcyclohexane	50.000	52.557	-5.1	99	0.00
40 TM	Benzene	50.000	50.584	-1.2	100	0.00
41 T	Methacrylonitrile	50.000	47.430	5.1	92	0.00
42 TM	1,2-Dichloroethane	50.000	50.924	-1.8	99	0.00
43 T	Isopropyl Acetate	50.000	48.023	4.0	95	0.00
44 TM	Trichloroethene	50.000	52.313	-4.6	104	0.00
45 C	1,2-Dichloropropane	50.000	51.064	-2.1#	101	0.00
46 T	Dibromomethane	50.000	48.943	2.1	97	0.00
47 T	Bromodichloromethane	50.000	52.770	-5.5	100	0.00
48 T	Methyl methacrylate	50.000	54.418	-8.8	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
 Data File : VX024756.D
 Acq On : 12 Oct 2021 10:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 01:45:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	811.157	18.9	75	0.00
50 S	Toluene-d8	50.000	51.118	-2.2	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	269.595	-7.8	103	0.00
52 CM	Toluene	50.000	52.049	-4.1#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	52.645	-5.3	100	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.079	-6.2	101	0.00
55 T	1,1,2-Trichloroethane	50.000	52.437	-4.9	102	0.00
56 T	Ethyl methacrylate	50.000	45.865	8.3	87	0.00
57 T	1,3-Dichloropropane	50.000	50.840	-1.7	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.502	2.2	91	0.00
59 T	2-Hexanone	250.000	278.596	-11.4	104	0.00
60 T	Dibromochloromethane	50.000	48.472	3.1	99	0.00
61 T	1,2-Dibromoethane	50.000	52.077	-4.2	100	0.00
62 S	4-Bromofluorobenzene	50.000	52.180	-4.4	103	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	57.133	-14.3	113	0.00
65 PM	Chlorobenzene	50.000	51.204	-2.4	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.399	-4.8	102	0.00
67 C	Ethyl Benzene	50.000	52.398	-4.8#	101	0.00
68 T	m/p-Xylenes	100.000	104.573	-4.6	99	0.00
69 T	o-Xylene	50.000	50.517	-1.0	97	0.00
70 T	Styrene	50.000	49.906	0.2	95	0.00
71 P	Bromoform	50.000	44.343	11.3	96	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	50.335	-0.7	100	0.00
74 T	N-ethyl acetate	50.000	47.280	5.4	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.225	7.5	95	0.00
76 T	1,2,3-Trichloropropane	50.000	49.218	1.6	100	0.00
77 T	Bromobenzene	50.000	49.290	1.4	102	0.00
78 T	n-propylbenzene	50.000	50.276	-0.6	100	0.00
79 T	2-Chlorotoluene	50.000	48.638	2.7	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.429	-4.9	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.299	7.4	96	0.00
82 T	4-Chlorotoluene	50.000	49.876	0.2	100	0.00
83 T	tert-Butylbenzene	50.000	49.660	0.7	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.279	-6.6	103	0.00
85 T	sec-Butylbenzene	50.000	48.780	2.4	94	0.00
86 T	p-Isopropyltoluene	50.000	53.295	-6.6	102	0.00
87 T	1,3-Dichlorobenzene	50.000	47.772	4.5	98	0.00
88 T	1,4-Dichlorobenzene	50.000	46.609	6.8	98	0.00
89 T	n-Butylbenzene	50.000	53.731	-7.5	104	0.00
90 T	Hexachloroethane	50.000	49.679	0.6	99	0.00
91 T	1,2-Dichlorobenzene	50.000	47.263	5.5	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.440	7.1	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.907	0.2	101	0.00
94 T	Hexachlorobutadiene	50.000	50.730	-1.5	102	0.00
95 T	Naphthalene	50.000	51.539	-3.1	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
Data File : VX024756.D
Acq On : 12 Oct 2021 10:19
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 13 01:45:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.402	1.2	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6