

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
 Data File : VX024765.D
 Acq On : 12 Oct 2021 16:01
 Operator : JC/MD
 Sample : M4169-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 12 16:48:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218968	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83600	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96138	54.220	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.440%
35) Dibromofluoromethane	5.397	113	74815	51.431	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.860%
50) Toluene-d8	8.653	98	269626	50.360	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.720%
62) 4-Bromofluorobenzene	11.085	95	101573	48.252	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.500%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101221\
Data File : VX024765.D
Acq On : 12 Oct 2021 16:01
Operator : JC/MD
Sample : M4169-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 12 16:48:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK


Tgt Ion:168 Resp: 125808
Ion Ratio Lower Upper
168 100
99 59.2 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 54.220 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01

Tgt Ion: 65 Resp: 96138
Ion Ratio Lower Upper
65 100
67 50.5 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK


Tgt Ion:114 Resp: 218968
Ion Ratio Lower Upper
114 100
63 20.7 0.0 40.8
88 16.2 0.0 32.8



#35
Dibromofluoromethane
Concen: 51.431 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01

Tgt Ion:113 Resp: 74815
Ion Ratio Lower Upper
113 100
111 100.7 81.9 122.9
192 19.6 16.7 25.1









#50
Toluene-d8
Concen: 50.360 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK


Tgt Ion: 98 Resp: 269626
Ion Ratio Lower Upper
98 100
100 62.6 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 48.252 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01

Tgt Ion: 95 Resp: 101573
Ion Ratio Lower Upper
95 100
174 80.7 0.0 154.6
176 77.5 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Tgt Ion:117 Resp: 208035
Ion Ratio Lower Upper
117 100
82 58.5 47.4 71.2
119 31.2 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX024765.D
Acq: 12 Oct 2021 16:01

Tgt Ion:152 Resp: 83600
Ion Ratio Lower Upper
152 100
115 61.3 41.1 123.3
150 157.8 0.0 349.0

