

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
 Data File : VX032079.D
 Acq On : 12 Oct 2022 13:12
 Operator : JC/MD
 Sample : N5100-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-TB-20221007

Quant Time: Oct 13 09:13:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	80100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121516	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124639	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87810	53.082	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.160%	
35) Dibromofluoromethane	5.385	113	65712	47.773	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.540%	
50) Toluene-d8	8.647	98	247673	50.040	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.080%	
62) 4-Bromofluorobenzene	11.079	95	82937	47.398	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.800%	
Target Compounds						
					Qvalue	
30) Chloroform	5.086	83	4333	1.663	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
Data File : VX032079.D
Acq On : 12 Oct 2022 13:12
Operator : JC/MD
Sample : N5100-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-TB-20221007

Quant Time: Oct 13 09:13:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

