

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
 Data File : VX032080.D
 Acq On : 12 Oct 2022 13:35
 Operator : JC/MD
 Sample : N5100-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-EB-20221010

Quant Time: Oct 13 09:13:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	120643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75438	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88816	53.392	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.780%	
35) Dibromofluoromethane	5.385	113	66292	48.544	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.080%	
50) Toluene-d8	8.647	98	248039	50.477	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.960%	
62) 4-Bromofluorobenzene	11.079	95	84819	48.825	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.640%	
Target Compounds						
30) Chloroform	5.093	83	4318	1.648	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
Data File : VX032080.D
Acq On : 12 Oct 2022 13:35
Operator : JC/MD
Sample : N5100-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-EB-20221010

Quant Time: Oct 13 09:13:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

