

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
 Data File : VX032082.D
 Acq On : 12 Oct 2022 14:22
 Operator : JC/MD
 Sample : N5106-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-101122

Quant Time: Oct 13 09:14:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	114067	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120140	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83487	52.786	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.580%
35) Dibromofluoromethane	5.385	113	62460	48.375	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.740%
50) Toluene-d8	8.653	98	235331	50.651	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.300%
62) 4-Bromofluorobenzene	11.079	95	79841	48.609	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.220%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	4469	8.314	ug/l	97
20) Methylene Chloride	2.788	84	2628	1.860	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
Data File : VX032082.D
Acq On : 12 Oct 2022 14:22
Operator : JC/MD
Sample : N5106-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-101122

Quant Time: Oct 13 09:14:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

