

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
 Data File : VX032083.D
 Acq On : 12 Oct 2022 14:45
 Operator : JC/MD
 Sample : N5106-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-01-312-332-101122

Quant Time: Oct 13 09:14:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	116478	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83877	52.781	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.560%	
35) Dibromofluoromethane	5.385	113	63496	48.159	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.320%	
50) Toluene-d8	8.647	98	240471	50.686	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.380%	
62) 4-Bromofluorobenzene	11.079	95	80748	48.143	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.280%	
Target Compounds						Qvalue
16) Acetone	2.386	43	2284	4.229	ug/l	98
52) Toluene	8.720	92	111988	33.456	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
Data File : VX032083.D
Acq On : 12 Oct 2022 14:45
Operator : JC/MD
Sample : N5106-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-01-312-332-101122

Quant Time: Oct 13 09:14:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

