

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
 Data File : VX032084.D
 Acq On : 12 Oct 2022 15:09
 Operator : JC/MD
 Sample : N5042-03 500X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EOR-0615

Quant Time: Oct 13 09:15:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	80481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118999	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88105	53.008	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.020%
35) Dibromofluoromethane	5.385	113	65561	48.672	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.340%
50) Toluene-d8	8.647	98	243868	50.313	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.079	95	83059	48.472	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.940%
Target Compounds						
16) Acetone	2.386	43	146858	259.978	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
Data File : VX032084.D
Acq On : 12 Oct 2022 15:09
Operator : JC/MD
Sample : N5042-03 500X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EOR-0615

Quant Time: Oct 13 09:15:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

